



SMILLIE McADAMS SUMMERLIN LTD.

900 MCKAY ROAD, UNIT 1, PICKERING, ONTARIO, CANADA L1W 3X8 • Phone: (905) 428-6900 • Fax: (905) 428-6598

SMS PH PROBE MAINTENANCE INSTRUCTIONS

INSPECT PH ELECTRODE (MODEL 9037) AND FLUSH WITH CLEAN WATER TO REMOVE ANY SILT OR DEBRIS.

A PH ELECTRODE WILL PROVIDE ACCURATE AND QUICK RESPONSE READINGS FOR A PERIOD OF 12-18 MONTHS. TO ENSURE A SUITABLE SIGNAL TO THE MONITORING PANEL (MODEL 47D), SMS RECOMMENDS SCHEDULED REPLACEMENT OF ELECTRODE ON AN ANNUAL BASIS.

FOR MORE INFORMATION AND TO ORDER PH PROBE CONTACT SMS LTD. PH 905-428-6900, FAX 905-428-6598

NOTE: MOST MUNICIPALITIES ALLOW WASTE WITH A PH BETWEEN APPROXIMATELY 5.5 AND 10. THIS PANEL IS SET TO ALARM AT A PH OF 5.